

Instructions for use & technical data CobraBond K



Manufacturer	Whitepeaks Dental Solutions GmbH & Co. KG Langeheide 9 - 45239 Essen - Germany
Product/ Product type	Metal alloy blanks for the production of individual dental restorations
Product form	Metal blanks in different diameter and thickness
Material type	Cobalt/ chrome alloy (type 4) – medical device class IIa
Circle of users	Instructed users who produce individual dental restorations

Indication/ intended use

CobraBond K is exclusively suitable for the production of dental products.

Indication

- anatomical reduced copings and pontics in anterior and posterior area
- bridges up to 14 units or bridges with small diameters
- primary and secondary telescopic crowns
- clasps, bars and retention constructions
- full anatomical crowns and bridges in anterior and posterior area
- restorations with small diameters which are exposed to high forces
- free end bridge constructions with maximum 1 pontic
- supra constructions for implant cases
- removable prosthesis

CobraBond K is a type 4 co/cr alloy.

Contraindication

Do not use in case of proven hypersensitivity against the alloy or one of its components.

Veneer ceramics

Co/Cr veneering porcelain

Material properties/ technical data

Composition:

Co	59,0% - 63,0%
Cr	27,0% - 29,0%
W	8,0% - 9,0%
Mn	0,15% - 0,35%
Fe	0,0% - 0,5%
Si	1,5% - 1,8%
C	0,0% - 0,1%

Technical data:

yield strength 0,2%	~ 360 – 484 MPa
elongation at break in percent	> 2 - 16 %
tensile strength	~ 400 - 900 MPa
density	~ 8 – 9 g /cm ³
corrosion resistance	< 200 µg cm ²
tarnish resistance	yes
Vickers hardness	~ 270 - 450 HV 10/30
coefficient of thermal expansion	14,1 x 10 ⁻⁶ /K

Specification

CobraBond K is a nickel- and beryllium free cobalt/ chrome blank, specially designed for CAD/CAM applications. The material is very homogenous and easy to mill, also it is suitable for laser welding.

Instructions for use

Removal of frameworks

Cut out and smoothen the surfaces of milled frameworks with carbide burs or separating discs. Please use the same cutter for one alloy to avoid contamination.

Veneering with ceramic

Basically all commercial veneering porcelains can be used. Please follow the instructions for use of your chosen veneering porcelain manufacturer and the coefficient of thermal expansion specified therein for compatibility.

The minimum thickness of the prepared coping should not be less than 0.3 mm. It's recommended to sandblast the frames with minimum 110 µm of aluminium oxide with 3-4 bar and clean with steam cleaner. Oxide firing is not mandatory but can be done as an option for 5 minutes at 980 °C with vacuum (cleaning firing). The frame needs to be sandblasted with aluminium oxide with about 110 µm and 3-4 bar to remove the present oxide layer thoroughly. In the end the cleaning by steam cleaner is mandatory. If you use a ceramic bonder please consider the instructions for use of the manufacturer.

Soldering

We recommend a chrome cobalt soldering metal for soldering. CopraBond K frames should not be soldered with gold or palladium solders. CopraBond K is easy to weld with a dental laser.

Safety instructions

Warning: Contains cobalt (Co). The dust produced during processing of this product may cause cancer, may reduce fertility and may probably cause genetic defects. Inhalation may cause allergies, asthma-like symptoms or breathing difficulties. Avoid inhalation/ contact with skin/ contact with eyes. Always wear respiratory protection (filter class FFP3), tightly fitting safety goggles, protective gloves and protective clothing and always switch on suction equipment with filter class Hepa H.

Storage

No special storage conditions. Store in the original packaging.

Disposal

Dispose of product and packaging in accordance with local/ regional/ national/ international regulations. Do not dispose of together with household waste. Do not allow to enter water, ground water or sewage system.

Notice

Any serious incident, that has occurred in relation to the device must be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is established.

Explanation of the markings on the packaging



Symbol for „article number“



Symbol for „LOT number“



Confirmation: The product complies with the applicable European directives.



Symbol for „number of products in package“



Symbol for „follow the instructions for use“

**Instructions for use
& technical data
CopraBond K**



Symbol for „is a medical device“



Symbol for “production date”

Rx only

Symbol for “Caution: US Federal law restricts this device to sale by or on the order of a licensed physician or dentist.”